



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 3 ml		
Product Code	P1006-F3ML	Batch no	2021050002
Material	PE	Packing	500 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 13mm	Caliper	Qualified
	Approximates OD of pipette: 8mm		
	Approximates of Length of bubble: 42mm	Caliper	Qualified
	Approximates overall Length: 160mm	Caliper	Qualified
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

**Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 1 ml		
Product Code	P1006-E1ML	Batch no	2021070009
Material	PE	Packing	100 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 12.5mm	Caliper	Qualified
	Approximates OD of pipette: 5mm		
	Approximates of Length of bubble: 40mm	Caliper	Qualified
	Approximates overall Length: 145mm	Caliper	Qualified
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

**Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 1 ml		
Product Code	P1006-F1ML	Batch no	2021070010
Material	PE	Packing	500 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 12.5mm	Caliper	Qualified
	Approximates OD of pipette: 5mm		
	Approximates of Length of bubble: 40mm	Caliper	Qualified
	Approximates overall Length: 145mm	Caliper	Qualified
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

**Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 10 ml		
Product Code	P1006-E10ML	Batch no	2021070011
Material	PE	Packing	100 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 17mm	Caliper	Qualified
	Approximates OD of pipette: 10mm	Caliper	Qualified
	Approximates of Length of bubble: 60mm	Caliper	Qualified
	Approximates overall Length: 280mm	Caliper	Qualified
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

**Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 3 ml,type A		
Product Code	P1006-E3ML	Batch no	2022010001
Material	PE	Packing	100 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 12mm	Caliper	Qualified
	Approximates OD of pipette: 7.5mm	Caliper	Qualified
	Approximates of Length of bubble: 41mm	Caliper	Qualified
	Approximates of tube below bubble length: 117mm	Caliper	Qualified
	Approximates overall Length: 158mm		
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

### **Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 3 ml,type A		
Product Code	P1006-F3ML	Batch no	2022010002
Material	PE	Packing	500 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 12mm	Caliper	Qualified
	Approximates OD of pipette: 7.5mm	Caliper	Qualified
	Approximates of Length of bubble: 41mm	Caliper	Qualified
	Approximates of tube below bubble length: 117mm	Caliper	Qualified
	Approximates overall Length: 158mm		
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

### **Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.



## Certificate of analysis

Product Description	Pasteur Pipette made of polyethylene 3 ml		
Product Code	P1006-E3ML	Batch no	2021050001
Material	PE	Packing	100 pcs/pk
Testing Items	Standard	Testing Method	Test result
Exterior	Appearance should be clean, bright, uniform color	Visual	Qualified
	Surface must not have foaming, deformation, crack	Visual	Qualified
	Bobble line with no width greater than 0.50mm	Visual	Qualified
	The scale should be clear, with obvious abrupt hand feeling	Visual	Qualified
Dimensions	Approximates OD of bubble: 13mm	Caliper	Qualified
	Approximates OD of pipette: 8mm		
	Approximates of Length of bubble: 42mm	Caliper	Qualified
	Approximates overall Length: 160mm	Caliper	Qualified
Physical property	The tube body is thin and flexible, with ideal liquid fluidity and good controllability	Visual	Qualified
	Leak: The container can not be leak when working under normal conditions	Visual	Qualified
Conclusion: the products are tested and meet the requirements.			

Remark: This product has been produced in clean room comply with Quality System Management Standard ISO 13485:2016

**Laboratory result: Satisfactory**

This document has been produced electronically and is valid without a signature.